

A.4 Sample calibration sheet

A typical calibration sheet example is shown in *Figure 55* below.

XMCT calibration sheet				
XMTC s/n	2630			
XMTC part number	XMTC-62-11			
Calibration part number	XMTC-CAL-311			
Calibration gases	H ₂ N ₂			
Calibration ranges	0 to 5%			
Work order number	508230006418			
Calibration date	April 25, 2009			
Technician	K. Brin			
Unit calibrated using	x10	Gain scale		
XMTC calibration data				
hydrogen in nitrogen				
	PT	%H ₂	x10 (mV)	Output (mA)
	1	0.00	-56.02	4.00
	2	5.00	574.00	20.00
Output	4 to 20 mA	0 to 5% H ₂ in N ₂		
Clamp output	No			
Field calibration				
Zero	100% N ₂			
Span	5% H ₂ in N ₂			

Figure 55: A sample calibration sheet